

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 9515 Accredited to ISO/IEC 17025:2017	National Manufacturing Institute Scotland (NMIS) part of University of Strathclyde	
	Issue No: 006 Issue date: 30 August 2022	Contact: Idil Temizyurek Tel: +44 (0) 141 534 5615 E-Mail: idil.temizyurek@strath.ac.uk Website: www.nmis.scot
85 Inchinnan Drive Inchinnan Renfrew PA4 9LJ		
Testing performed by the organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Location details		Technology Centre Activity	Location code
Location Address 85 Inchinnan Drive Inchinnan Renfrew PA4 9LJ	Local contact Idil Temizyurek Tel: +44 (0) 141 534 5615 E-Mail: idil.temizyurek@strath.ac.uk	Advanced Forming Research Centre - AFRC Residual Surface Stress	AFRC
Location Address 85 Inchinnan Drive Inchinnan Renfrew PA4 9LJ	Local contact Idil Temizyurek Tel: +44 (0) 141 534 5615 E-Mail: idil.temizyurek@strath.ac.uk	Digital Factory CMM Measurements	AFRC



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

National Manufacturing Institute Scotland (NMIS)
part of University of Strathclyde

Issue No: 006 Issue date: 30 August 2022

Testing performed by the organisation at the locations specified below

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location
Metallic Materials and Alloys / Formed and Forged Components, Machined Surfaces	<u>Residual Surface Stress</u> Residual Stress Measurements by X-Ray Diffraction (XRD)	NPL's Measurement Good Practice Guide No. 52 – Issue 2	AFRC
	Determination of Residual Stresses by the Hole-Drilling Strain-Gage Method	ASTM E837-20	AFRC
Metallic Materials and Alloys / Formed and Forged Components, Machined Surfaces	<u>CMM Measurements</u> General dimensional measurements, made using a coordinate measuring machine, with best measurement capability of:	Customer drawings and specifications. Laboratory procedures based on NPL Good Practice Guides 41 – Issue 2 and 43 – Issue 2	AFRC
	Length, Diameter and Position 0 to 900 x 900 x 700 mm – 5.9 + (21 x length in m) µm Angle – 2.0 minutes of arc		
END			